


**CAREGEN CO., LTD.**

46-38 Caregen Bldg., LS-ro 91beon-gil, Dongan-gu,
Anyang-si, Gyeonggi-do 14119 South Korea
TEL : +82-(0)31-420-9200 FAX : +82-(0)31-452-9299
Homepage : www.caregen.com

Certificate of Analysis

Product	DR.CYJ Hair Filler	Date of Issued	2018.07.31
Lot No.	DCHF-18-05	Quantity	22,562
Date of Manufacturing	2018.07.10	Date of Release	NA
Expiry Date	2020.07.09	Packaging Unit	1ml * 1ea

Test Items 試験項目	Specifications 規格	Test Result 試験結果	P/F
Appearance 外見	Viscous transparent Gel 粘性透明ゲル	Viscous transparent Gel	Pass
Volume 比重	$\geq 1.0 \text{ mL(g)}$	1.20 g	Pass
pH (25℃) pH値	7.3 ± 1.0	6.9	Pass
Assay :Sodium Hyaluronate ヒアルロン酸Na含有量	90.0 ~ 110.0 %	100.50%	Pass
Osmolarity 浸透圧	270 ~ 345 mOsmol/kg	302 mOsmol/kg	Pass
Complex Viscosity (1Hz, 25℃) 粘度	0.1~5.0 Pa.s	0.5046 Pa.s	Pass
Bacterial endotoxins エンドトキシン試験	$\leq 12.5 \text{ EU/mL}$	2.90EU/ml	Pass
Sterility 無菌試験	No growth	No growth	Pass

Package & Storage	The product is sensitive to a long term of high temperature and strong radiation. Product is stable at room temperature and the storage at 4℃ is recommended for long-term preservation. The product is packaged in a sterilized syringe.
Note	N.A / Store between 2℃ and 25℃, away from sunlight. Make sure there are no visible signs of damage to the packaging before use.
<div>Reported by <u>QC Manager</u> / Date <u>2018. 7. 31</u></div> <div>Approved by <u>QMR</u> / Date _____</div> <div></div>	